


Bias=3.62 (25.17%) MAE=3.62 $r^2=1.00$ Precision=3.62 (25.21%)


PM10 SDS011 14595225



PM10 reference EDM164 OPC004

Bias=-0.32 (-5.73%) MAE=0.32 $r^2=1.00$ Precision=0.33 (5.99%)


PM2.5 SDS011 14595225



PM2.5 reference EDM164 OPC004

Bias=0.15 (1.86%) MAE=0.31 $r^2=0.71$ Precision=0.40 (5.06%)

Temp BME280 14595225



Temp reference SHT 177410813

Bias=-1.79 (-2.34%) MAE=4.05 $r^2=0.49$ Precision=4.53 (5.94%)

